



Australian Government
Department of Health and Aged Care
Therapeutic Goods Administration

Public Summary

Summary for ARTG Entry: 158581 Dynek Pty Ltd - Catgut Plain - Suture, catgut

ARTG entry for Medical Device Included Class III
Sponsor Dynek Pty Ltd
Postal Address PO Box 2346, PORT ADELAIDE, SA, 5015
Australia
ARTG Start Date 19/01/2009
Product Category Medical Device Class III
Status Active
Approval Area Medical Devices

Conditions

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

Manufacturers

Name	Address
Dynek Pty Ltd	9 Circuit Drive HENDON, SA, 5014 Australia

Products

1 . Catgut Plain - Suture, catgut

Product Type	Single Device Product	Effective Date	19/01/2009
GMDN	13898 Suture, catgut		
Functional Description	Suture, absorbable, non-synthetic, pseudo-monofilament, with surgeon able to choose appropriate thread diameter and length according to surgical requirement		
Intended Purpose	Surgical procedures for tying off, ligation and/or tissue approximation		
Variant information	Length (cm) 10 - 250 Diameter (mm) 0.070 - 0.899		

Specific Conditions

No Specific Conditions included on Record

© Commonwealth of Australia. This work is copyright. You are not permitted to re-transmit, distribute or commercialise the material without obtaining prior written approval from the Commonwealth. Further details can be found at <http://www.tga.gov.au/about/website-copyright.htm>.

Public Summary